

This set up is designed to study coriolis component of Acceleration of a slider crank Mechanism. The apparatus uses hydraulic analogy to represent the rotating slider. It consists of a rotating block with two arms in opposite direction. These tubes can be rotated at various speeds by using a swinging field motor, which also acts as a dynamometer to measure torque applied to rotating tubes. A Perspex window on top cover helps to visualize the process. Rotameter is used to measure water flow rate through tubes. Water is circulated by small monoblock pump.

Specifications:

1. Main Tank with fiberglass lining.
2. Rotating Arms 9/6 mm dia, 300 mm long.
3. Motor - Swinging field, D.C., 0.5 HP.
4. Rotameter
5. Monoblock Pump
6. Control Panel comprising of -
 - (i) Speed Control Unit.
 - (ii) Speed Indicator.
 - (iii) Necessary switches.
7. Rigid support frame.

Range of Experiment:

Coriolis Component of Acceleration can be determined at various speeds of rotation and water flow rates.

Service Required:

1. A.C. Single Phase .230 V. stabilized supply.
2. Floor Space -1.5 m X 1.5 m.

Note: Specifications are subject to change.

Tesca Technologies Pvt. Ltd.

IT-2013, Ramchandrapura Industrial Area, Sitapura Extension,
Near Bombay Hospital, Vidhani Circle, Jaipur-302022, Rajasthan, India,
Tel: +91-141-2771791 / 2771792; Email: info@tesca.in, tesca.technologies@gmail.com
Website: www.tesca.in

