



The test rig comprises of a two-stage air compressor fitted with an air tank with orifice, pressure gauges and an energy meter to measure the input & a Temperature indicator measures the temperatures at various points. The compressor is mounted on air receiver along with the motor, and is provided with air relief valve (safety) and a pressure switch. The unit determines volumetric efficiency and isothermal efficiency at various discharge pressures

SPECIFICATIONS:

1. Air compressor - Double cylinder, two-stage type driven by a 2-hp. Three - phase motor mounted on air receiver provided with delivery valve.
2. Air tank and orifice with water manometer for air intake measurement.
3. Pressure gauges at outlet on both stages.
4. Digital temperature indicator.
5. Energy meter to measure input power. 6. A technical manual accompanies the equipments.

SERVICES REQUIRED:

1. Floor space of 2,5 X 1.5m.
2. 440 V, 15A, 3ph. AC supply with neutral and earthing connection.
3. A hand tachometer.

Note: Specifications are subject to change.

Tesca Technologies Pvt. Ltd.

IT-2013, Ramchandrapura Industrial Area, Sitapura Extension,
Near Bombay Hospital, Vidhani Circle, Jaipur-302022, Rajasthan, India,
Tel: +91-141-2771791 / 2771792; Email: info@tesca.in, tesca.technologies@gmail.com
Website: www.tesca.in